CertaCal
CALIBRATION MANAGEMENT

Demo Scale Services
1 Example Street
Springfield, CA 90000
(555) 010-0100

Certificate of Calibration

CERTIFICATE

Certificate No.	DEMO-2026-000001
Job / CSR No.	—
Calibration date	11/18/2025
Next due	11/09/2026

CUSTOMER

Customer	Sunrise Produce Co. (demo)
Location	100 Orchard Way Springfield, CA 90000
Unit location (site)	Packing House
Customer ID	—

SCALE

Manufacturer	Rice Lake
Model	M-BS101
Serial number	BS-2201
Capacity	60
Division	0.02
Divisions	3000
Tested as	—
Platform manufacturer	
Platform model	
Platform serial number	

CALIBRATION DETAILS

Calibration date	11/18/2025
Procedure	Bench Scale calibration
Work instruction	CP-100 Scale Calibration: Overview and Order of Work
Technician	Terry Tech (demo)

ENVIRONMENTAL CONDITIONS

Temperature	22 °C
Humidity	40 %

VISUAL INSPECTION

Not recorded.

Result: IN TOLERANCE

SHIFT TEST

N/A — see Notes below

MEASUREMENT RESULTS

#	NOMINAL	AS FOUND	AS LEFT	UNIT	UNCERTAINTY	TOLERANCE	RESULT
1	0	0.00	0.00	lb	0.045	±0.02	PASS
2	30	30.00	30.00	lb	0.045	±0.02	PASS
3	60	60.00	60.00	lb	0.045	±0.02	PASS

NOTES

Calibrated and found within tolerance.

METROLOGICAL TRACEABILITY

Measurements are traceable to the SI through NIST by way of these reference standards:

Standard	Certificate	Traceable to
	—	

AUTHORIZED BY

Dana Office (demo), Owner
Technician: Terry Tech (demo)

Technician's signature for the job: Terry Tech (demo), 11/18/2025

Acknowledged by Pat Rivera (demo): "I acknowledge that the equipment listed was serviced on this date and the results were reviewed with me."

Verify this certificate: https://demo.certacal.net/verify/c/39d7e2b5-eb26-4b57-82a2-f803ef0ff40f/

DEMO - NOT A CALIBRATION CERTIFICATE

Decision rule: pass/fail is decided by simple acceptance (ILAC G8): a reading passes when its error is within the tolerance shown. Measurement uncertainty is not taken into account.

Uncertainty is the expanded uncertainty U: the combined standard uncertainty multiplied by a coverage factor k = 2, giving a coverage probability of approximately 95 %.

- Tested in accordance with NIST Handbook 44 (demo).

Demo Scale Services — this certificate relates only to the item identified above and shall not be reproduced except in full, without written approval.

DEMO - NOT A CALIBRATION CERTIFICATE